TEST REPORT

REPORT NO.: CTNT2609150190101

Product name:	AMIRO Spectra 10-in-1 LED Light Therapy Mask
Model No.:	L3
Applicant:	Shenzhen Accompany Technology Co., Ltd.
Test procedure:	Entrustment Test

Shenzhen Zhongwei Testing Technology Co., Ltd.

This report may not be reproduced in part without permission to avoid ambiguous interpretation.
This report can be checked and verified in the following ways.
Tel：0755-28680489 E-mail：admin@ctnt-cert.com Web：www.ctnt-cert.com

TEST REPORT

IEC 62471:2006
Photobiological safety of lamps and lamp systems

Report Number.....:	CTNT2609150190101
Date of issue.....:	Oct. 23, 2026
Name of Testing Laboratory preparing the Report.....:	Shenzhen Zhongwei Testing Technology Co., Ltd. Room 1A106 & 2C230, No.109, Lijia Road, Henggang, Henggang Subdistrict, Longgang District, Shenzhen, Guangdong, China Tel: 086-13662574496 E-mail: admin@ctnt-cert.com Web: www.ctnt-cert.com
Applicant's name.....:	Shenzhen Accompany Technology Co., Ltd.
Address.....:	25th Floor, Gaoxin Community United Headquarters Building, No.63, Gaoxin South 10th Road, Binhai Community, Yuehai Street, Nanshan District, Shenzhen, China
Manufacturer's name.....:	Shenzhen Accompany Technology Co., Ltd.
Address.....:	25th Floor, Gaoxin Community United Headquarters Building, No.63, Gaoxin South 10th Road, Binhai Community, Yuehai Street, Nanshan District, Shenzhen, China
Test specification:	
Standard.....:	IEC 62471:2006
Test procedure.....:	Entrustment Test
Non-standard test method.....:	N/A
Test Report Form No. .....:	IEC 62471B
Test Report Form(s) Originator ....:	VDE Testing and Certification Institute
Master TRF.....:	Dated 2018-08-16

General disclaimer:

The test results presented in this report relate only to the object tested.

This report shall not be reproduced, except in full, without the written approval of the Issuing CTNT Testing Laboratory.

The authenticity of this Test Report and its contents can be verified by contacting the CTNT, responsible for this Test Report.

Test item description.....:	AMIRO Spectra 10-in-1 LED Light Therapy Mask
Trade Mark.....:	N/A
Model/Type reference.....:	L3

Test item particulars	
Tested lamp	N/A
Tested lamp system.....	☒ continuous wave lamps ☐ pulsed lamps
Lamp classification group.....	☒ exempt ☐ risk 1 ☐ risk 2 ☐ risk 3
Lamp cap	N/A
Bulb	Non-replaceable LED Bulb
Rated of the lamp.....	5V⎓ 2A 10W
Furthermore, marking on the lamp	N/A
Seasoning of lamps according IEC standard.....	N/A
Used measurement instrument.....	Photobiological safety test system
Temperature by measurement.....	24.2°C
Information for safety use.....	N/A

Responsible Testing Laboratory (as applicable), testing procedure and testing location(s):

Laboratory Name	Shenzhen Zhongwei Testing Technology Co., Ltd.	
Testing location/ address.....:	Room 114, 2/F., Building C, No.108, Lijia Road, Henggang Street, Longgang District, Shenzhen, Guangdong, China	
Prepared by(Reporting Engineer) .....:	Brandon Li	Brandon Li
Reviewed by(Supervisor).....:	Oliver Long	Oliver Long
Approved by(Chief Engineer) .....:	Flight Lee	

List of Attachments (including a total number of pages in each attachment):

- Photo document
- Test data graph

Summary of testing:

Tests performed (name of test and test clause):	**Testing location:**
Determination of the result includes consideration of measurement uncertainty from the test equipment and methods. According to the test of IEC 62471:2006 Photobiological safety of lamps and lamp systems. **The classification of this product is in the Exempt Group.**	Shenzhen Zhongwei Testing Technology Co., Ltd. Room 114, 2/F., Building C, No.108, Lijia Road, Henggang Street, Longgang District, Shenzhen, Guangdong, China Tel: 086-755-28680489 E-mail: admin@ctnt-cert.com Web: www.ctnt-cert.com

This report may not be reproduced in part without permission to avoid ambiguous interpretation.
This report can be checked and verified in the following ways.
Tel：0755-28680489 E-mail：admin@ctnt-cert.com Web：www.ctnt-cert.com